

ICS 27.100
CCS F 24

DL

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

DL/T 356-2010

**Calibration specification of partial discharge measuring
devices**

(English Translation)

Issued on: May 24, 2010

Implemented on: October 1, 2010

Issued by: National Energy Administration

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the partial discharge side of the meter (referred to PD meter) the general technical requirements, measuring performance requirements, calibration methods and calibration results processing. This standard applies to using pulse current law, the frequency range 10kHz ~ 500kHz, bandwidth 40kHz ~ 400kHz PD instrument calibration.

Document data

International classification (ICS): 27.100.

Chinese classification (CCS): F24.

Estimated length of the Chinese original: 21,281 words.

Standards quoted by this document

GB/T 7354-2003; GB/T 16927.1