

ICS 27.100
CCS K 20

DL

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

DL/T 677-2018

Replacing DL/T 677-2009

**Inspection code of on-line chemical instruments for power
plant**

发电厂在线化学仪表检验规程
(English Translation)

Issued on: April 3, 2018

Implemented on: July 1, 2018

Issued by: National Energy Administration

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the technical requirements, test conditions and test methods for on-line conductivity, pH, sodium ion, dissolved oxygen and silicate instruments of power plants. This standard is applicable to the acceptance inspection and the measurement inspection during the operation of the above-mentioned online chemical instrument in the power plant. Laboratory chemical instruments can refer to this standard.

Document data

International classification (ICS): 27.100.

Chinese classification (CCS): K20.

Estimated length of the Chinese original: 48,471 words.

Supersedes the earlier edition DL/T 677-2009.

Standards quoted by this document

GB/T 6903; GB/T 12148; GB/T 12149; GB/T 13966; GB/T 27501; DL/T 913; JJG 119; JJG 291; JJG 376; JJG 757