

ICS 11.140
CCS G45



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB 24788-2025

Replacing GB 24788-2009

Safety technical requirements for medical gloves

医用手套安全技术要求

Issued on: May 30, 2025

Implemented on: June 1, 2028

**Issued by: State Administration for Market Regulation;
Standardization Administration of China**

Table of Contents

2 Normative References	
3 Terms and Definitions	
4 Safety Technical Requirements	4.1 Limit for Water-extractable Protein
4.3 Microbiological Indicators	

Foreword

NATIONAL STANDARD OF THE PEOPLE'S REPUBLIC OF CHINA ICS

Replacing GB 24788-2009

Safety Technical Requirements for Medical Gloves

7 Table of Contents Foreword	3
1 Scope	4
2 Normative References	4
3 Terms and Definitions	4
4 Safety Technical Requirements	5

Safety Technical Requirements for Medical Gloves 1 Scope

This document specifies the safety technical requirements for medical gloves. This document applies to disposable medical rubber examination gloves, disposable sterile rubber surgical gloves, disposable non-sterile rubber surgical gloves, and disposable polyvinyl chloride medical examination gloves.

2 Normative References

The contents of the following documents constitute indispensable clauses of this document through the normative references in the text. In terms of references with a specified date, only versions with a specified date are applicable to this document. In terms of references without a specified date, the latest version (including all the modifications) is applicable to this document.

GB/T 14233.1-2022 Test Methods for Infusion, Transfusion, Injection Equipment for Medical Use - Part 1: Chemical Analysis Methods

GB 15979-2024 Hygienic Requirements for Disposable Sanitary Products

GB/T 21869-2008 Medical Gloves - Determination of Removable Surface Powder

GB/T 21870-2008 Medical Gloves Made from Natural Rubber Latex - Determination of Water-extractable Protein Using the Modified Lowry Method Pharmacopoeia of the People's Republic of China (Version 2020) Volume IV

3 Terms and Definitions

The following terms and definitions are applicable to this document.

3.1 powder Under test conditions, any water-insoluble adhering matter on the surface of medical gloves that can be removed by washing with water.

3.2 powdered medical gloves Medical gloves that are typically produced through the powder coating process during the production process for ease of donning.

7 Medical gloves produced in the glove production process where the powder coating process is intentionally not used, or where the powder coating process is used, but the power is intentionally removed.

3.4 water-extractable protein Proteins and polypeptides that exist in natural rubber latex products and can be extracted with water.

4 Safety Technical Requirements 4.1 Limit for Water-extractable Protein

Medical gloves made from natural rubber latex shall contain no more than 200 g/dm² of water-extractable protein.

4.2 Limit for Residual Surface Powder The residual surface powder content of powdered medical gloves shall not exceed 15 mg/dm²

2. The residual surface powder content of powder-free medical gloves shall not exceed

2.0 mg/piece. > NOTE: the limit for residual surface powder does not apply to disposable polyvinyl chloride medical examination gloves.

4.3 Microbiological Indicators

4.3.1 The microbiological indicators of non-sterile medical gloves shall comply with the provisions of Table 1.

4.3.2 Sterilized gloves shall be sterile.

--- The End ---