

ICS 17.220.20
CCS N 25



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 11151-1989

Alternating current bridge

交流电桥

(English Translation)

Issued on: March 31, 1989

Implemented on: April 1, 1990

Issued by: SAMR; SAC

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the technical requirements of AC bridge, test methods, inspection rules and signs, packaging requirements. This standard applies to the measured capacitance C ($10^{-10} \sim 10^4$ F), the inductance L ($10^{-9} \sim 10^3$ H), the resistance R ($10^{-2} \sim 10^7$ Ω), loss tangent $\tan \delta$ (10^{-5} to 2), the quality factor Q (0.5 ~ 500), conductance G ($1 \sim 10^7$ S) AC balanced bridge. This standard does not apply to automatic and semi-automatic bridge bridge and special measurement used.

Document data

International classification (ICS): 17.220.20.

Chinese classification (CCS): N25.

Estimated length of the Chinese original: 1,170 words.

Standards quoted by this document

GB 6738; ZB Y002; ZB Y003