

ICS 13.060.40
CCS Z 16



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 11904-1989

**Water quality. Determination of potassium and sodium. Flame
atomic absorption spectrophotometry
(English Translation)**

Issued on: November 3, 1989

Implemented on: July 1, 1990

Issued by: State Bureau of Technical Supervision

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the use of flame atomic absorption spectrophotometry can be filtered state of potassium and sodium. He applied to the determination of surface water and drinking water. Potassium measurement range 0.05 ~ 4.00mg/L; sodium was 0.01 ~ 2.00mg/L. For higher concentrations of potassium and sodium samples should be taken fewer samples for analysis, or the use of time-sensitive line measurement.

Document data

International classification (ICS): 13.060.40.

Chinese classification (CCS): Z16.

Estimated length of the Chinese original: 4,497 words.