

ICS 13.060.40

CCS Z 16



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 11911-1989

**Water quality. Determination of iron and manganese. Flame
atomic absorption spectrometric method**

水质 铁、锰的测定 火焰原子吸收分光光度法

(English Translation)

Issued on: November 3, 1989

Implemented on: July 1, 1990

Issued by: State Bureau of Technical Supervision

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard applies to surface water, groundwater and industrial wastewater iron, manganese measured. Iron, manganese and detection limits were 0.03mg/L and 0.01mg/L, the concentration range of the calibration curve was 0.1 ~ 5mg/L and 0.05 ~ 3mg/L.

Document data

International classification (ICS): 13.060.40.

Chinese classification (CCS): Z16.

Estimated length of the Chinese original: 4,463 words.