

ICS 81.040.01

CCS N64



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 12808-2015

Replacing GB/T 12808-1991

Laboratory glassware - One mark pipettes

实验室玻璃仪器 单标线吸量管

Issued on: December 31, 2015

Implemented on: July 1, 2016

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine of the People's Republic of China;
Standardization Administration of China**

Table of Contents

1	Scope
2	Normative References
3	Terms and Definitions
4	Specification Series and Accuracy Grades
5	Technical Requirements
5.1	Structure
6	Test Methods...
7	Inspection Rules...
8	Marking, Packaging, Transportation and Storage
8.1	Marking
8.2	Packaging

1 Scope

GB/T 12808-2015 is the Chinese national standard on laboratory glassware - one mark pipettes, in the field of glass and ceramics industries. The /T suffix marks it as a recommended standard: it is not compulsory by itself, but it becomes binding as soon as a contract, a tender or a customer specification calls it up - which in practice is how most foreign buyers meet it. It was issued on 31 December 2015 by the General Administration of Quality Supervision, Inspection and Quarantine of the People's Republic of China; Standardization Administration of China, and has been in force since 1 July 2016.

Classification: ICS 81.040.01, CCS N64. This page is published from the official record of the standard held by the Chinese standards administration: the identification, the dates, the classification and the issuing body are taken from there. The clause text, the tables and the numeric limits are in the document itself, which is delivered complete in English translation.

This Standard specifies the specification series, requirements, test methods, inspection rules, as well as marking, packaging, transportation and storage of one-mark pipette products used in ordinary laboratories. This Standard is applicable to the production and application of one-mark pipettes used in ordinary laboratories.

2 Normative References

The contents of the following documents constitute indispensable clauses of this document through the normative references in the text. In terms of references with a specified date, only versions with a specified date are applicable to this document. In terms of references

without a specified date, the latest version (including all the modifications) is applicable to this document.

GB/T 2828.1 Sampling Procedures for Inspection by Attributes - Part 1: Sampling Schemes Indexed by Acceptance Quality Limit (AQL) for Lot-by-lot Inspection

GB/T 6543 Single and Double Corrugated Boxes for Transport Packages

GB/T 6582 Glass - Hydrolytic Resistance of Glass Grains at 98 C - Method of Test and Classification

GB/T 15726 Glassware - Stress Examination Methods

GB/T 28209 Methods for Chemical Analysis of Borosilicate Glasses JJG 196 Working Glass Container

3 Terms and Definitions

The following terms and definitions are applicable to this Standard.

3.1 standard temperature Standard temperature is the temperature used when measuring in or out of its nominal volume (capacity), which is set to 20 C.

3.2 capacity At 20 C, the volume of water at 20 C contained when reaching the scale mark, expressed in (mL).

3.3 curved liquid surface setting Curved liquid surface setting means the lowest point of the curved liquid surface is tangent to the horizontal plane of the upper edge of the graduation line. The line of sight is on the same level as the upper edge of the graduation line.

4 Specification Series and Accuracy Grades

4.1 Product Specifications 1 mL, 2 mL, 3 mL, 5 mL, 10 mL, 15 mL, 20 mL, 25 mL, 50 mL and 100 mL.

4.2 Accuracy Grades Grade-A and Grade-B.

5.1 Structure

5.1.1 Structural composition The one-mark pipette consists of a straw, a reservoir and a liquid-flowing tube. These 3 parts are all straight-shaped and on the same axis. The structural composition of the one-mark pipette is shown in Figure 1. 1 mL and 2 mL Grade-B one-mark pipettes can be similar straight-shaped tubes or bubble-shaped tubes with a larger capacity. If they are bubble-shaped, they shall be approximately spherical and located between the pipettes and the nominal capacity line. The products of the same variety and specification, and delivered at the same time constitute one batch.

7.2.3 Inspection implementation and inspection results Sampling inspection shall be carried

out by the manufacturer in accordance with the exit-factory inspection items in Table 7. Only batch of products that are qualified in the inspection can be allowed to exit the factory, and a certificate of conformity shall be attached when exiting the factory.

8.1 Marking

8.1.1 Product marking The following markings shall be securely marked on each one-mark pipette:

- a) Manufacturer's trademark;
- b) Standard temperature "20 C";
- c) Figure of nominal capacity and capacity unit "mL";
- d) Grade symbol "A" or "B";
- e) Exit symbol "Ex".

8.1.2 Packaging marking The following markings shall be securely marked on the packaging box:

- a) There shall be pictorial markings for handling of goods "HANDLE WITH CARE" and "KEEP DRY";
- b) Product name, specifications, quantity, net weight, gross weight and volume;
- c) Manufacturer's name, registered trademark and production date;
- d) Manufacturer's address and telephone number.