

ICS 17.16  
CCS L 15



**NATIONAL STANDARD OF THE  
PEOPLE'S REPUBLIC OF CHINA**

**GB/T 13823.14-1995**

---

**Methods for the calibration of vibration and shock pick-ups  
Primary calibration by centrifuge  
(English Translation)**

**Issued on: April 2, 1995**

**Implemented on: February 1, 1996**

---

**Issued by: State Bureau of Technical Supervision**

## Table of Contents

|                                 |   |
|---------------------------------|---|
| What this standard covers ..... | 1 |
| Document data .....             | 2 |

### What this standard covers

This standard specifies the vibration and shock sensor calibration using a centrifuge equipment and calibration methods detailed requirements. This standard applies to strain, such as linear motion piezoresistive accelerometer has a zero-frequency response, including standards and working sensor. This standard applies to the zero-frequency acceleration:  $10 \sim 1000 \text{ m/s}^2$ . This standard applies to the error limit: reading  $\pm 1\%$ .

### Document data

International classification (ICS): 17.16.

Chinese classification (CCS): L15.

Estimated length of the Chinese original: 6,677 words.

Adopted international standard: ISO 5347-7-1993, MOD.