

ICS 17.140.01
CCS A 59



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 15611-1995

**Acoustics. Calibration of high frequency hydrophone
(English Translation)**

Issued on: July 3, 1995

Implemented on: February 1, 1996

Issued by: State Bureau of Technical Supervision

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the transducer with two calibrated reciprocity method of high frequency piezoelectric hydrophone approach. This standard applies to the frequency range of 0. 5 ~ 10MHz, the frequency range is calibrated using a high-frequency hydrophone 0. 1 ~ 0. 5MHz of GB/T 3223.

Document data

International classification (ICS): 17.140.01.

Chinese classification (CCS): A59.

Estimated length of the Chinese original: 8,810 words.

Adopted international standard: IEC 60866-1987, NEQ.