

ICS 17.16
CCS A 59



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 15951-1995

**Mechanical coupler for measurements on bone vibrators
(English Translation)**

Issued on: December 21, 1995

Implemented on: August 1, 1996

Issued by: State Bureau of Technical Supervision

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the requirements for power coupler. Power coupler in the frequency range of 125Hz ~ 8000Hz used to calibrate the bone conduction audiometry, measuring bone vibrator (bone conduction headphones) and measuring bone conduction hearing aids.

Document data

International classification (ICS): 17.16.

Chinese classification (CCS): A59.

Estimated length of the Chinese original: 10,144 words.

Adopted international standard: IEC 60373-1990, IDT.