

ICS 17.140.10
CCS A 59



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 16540-1996

**Acoustics. Measurement and characterisation of ultrasonic
fields in the frequency range 0.5MHz to 15MHz. Hydrophone
method**

(English Translation)

Issued on: September 13, 1996

Implemented on: March 1, 1997

Issued by: State Bureau of Technical Supervision

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the definition of 0.5 ~ 15MHz frequency range of internal water and other liquids in a variety of acoustic parameters of ultrasound, and has been calibrated using a piezoelectric hydrophone measurements for medical ultrasound equipment ultrasonic field parameters in water conditions arising and methods.

Document data

International classification (ICS): 17.140.10.

Chinese classification (CCS): A59.

Estimated length of the Chinese original: 31,359 words.

Adopted international standard: IEC 61102-1991, MOD; IEC 61102 AMD.1-1993, MOD.

Standards quoted by this document

GB/T 3947-1996; GB/T 4128-1995; GB/T 15611-1995; IEC 50 (801); IEC 469-1: 1987; IEC 854: 1986