

ICS 11.040.50

CCS C 43



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 17006.6-2003

**Evaluation and routine testing in medical imaging departments
-- Part 2-4: Constancy tests -- Hard copy cameras**

医用成像部门的评价及例行试验 第2-4部分：硬拷贝照相机稳定性试验

(English Translation)

Issued on: January 27, 2003

Implemented on: July 1, 2003

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine of the People Republic of China**

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard applies to: the monochrome continuous photosensitive material (such as optical and infrared radiation sensitive film material) generated image hard copy camera, and includes the use of a cathode ray tube, a laser beam, or thermoelectric printing system a variety of cameras, such as In diagnostic imaging systems applications -: Digital Radiography, Digital, count body layer helmet shadow imaging, magnetic resonance imaging, ultrasound imaging, nuclear medicine imaging. This test method is based on test pattern use basis *. This standard does not apply to the use of nuclear medicine-XY (analog) recording system. This standard is a special publication (standards and technical reports) Series part, described as GB/T 17006. 1-2000 (see Chapter 2), it gives diagnostic imaging performance stability test methods.

Document data

International classification (ICS): 11.040.50.

Chinese classification (CCS): C43.

Estimated length of the Chinese original: 20,287 words.

Adopted international standard: IEC 61223-2-4-1994; IDT.