

ICS 11.180.20
CCS C 45



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 17257.2-2009

Replacing GB/T 17257.2-1998; GB/T 17257.3-1998; GB/T 17257.4-1998

Urine collection bags -- Part 2: Requirements and test methods

集尿袋 第2部分：要求和检验方法

(English Translation)

Issued on: March 16, 2009

Implemented on: September 1, 2009

Issued by: Standardization Administration of China

Table of Contents

Introduction
1 Scope
2 Normative references
3 Terms and definitions
4 Requirements
5 General test conditions
6 Test Method
7 Inspection reports
Appendix A (normative) set the size of the bag

Foreword

GB/T 17257 "urine collection bag" is divided into two parts.

--- Part 1. Terminology

--- Part 2. Requirements and test methods this section GB/T Part of 217,257.

This section identical with ISO 8669-2:1996 "collection bags - Part 2. Requirements and test methods" (in English).

This Part replaces GB/T 17257.2-1998, GB/T 17257.3-1998, GB/T 17257.4-1998.

This part of GB/T 17257.2 ~ 17257.4-1998 compared to the main changes are as follows.

--- The GB/T 17257.2-1998, GB/T 17257.3-1998, GB/T 17257.4-1998 merger GB/T 17257 of this section. Wherein, GB/T 17257.2, GB/T 17257.3, GB/T 17257.4, respectively, as GB/T 17257 Part Chapter 4 points 4.9,4.1 and 4.2;

--- An increase of some terms and definitions;

--- Load increased leakage, the impact of the leak, the one-way flow valve reflux, attachment strength, introducing pipe strength requirements and the corresponding test methods.

This section of Appendix A normative appendix.

This section proposed by the People's Republic of China Ministry of Civil Affairs.

This part of the National Rehabilitation and special equipment Standardization Technical Committee on Persons with Disabilities (SAC/TC148) centralized.

This section drafted by. National Rehabilitation Device Quality Supervision and Inspection Center, China Rehabilitation Research Center, Beijing Poly Electronics Co., Ltd.
Hung-cheung.

1 Scope

GB/T 17257 in this section specifies the requirements and test methods for the following open-ended and closed-storage bag of.

- a) wore a urine collection bag (wearable);
- b) Hook suspension or floor frame urine collection bag (non-wearable).

This section does not apply to the urethra regeneration surgery, female and children with collection bags.

2 Normative references

The following documents contain provisions which, through reference GB/T 17257 in this section, constitute provisions of this part. For dated reference documents Member, all subsequent amendments (not including errata content) or revisions do not apply to this section, however, encouraged to reach under this section Parties to research agreement to use the latest versions of these documents. For undated reference documents, the latest versions apply to this section.

GB/T 17257.1 collection bags - Part 1. Terminology (GB/T 17257.1-2009, ISO 8669-1. 1988, IDT)

3 Terms and Definitions

GB/T 17257.1 and identified the following terms and definitions apply to this section.

Soft opening of the urine collection bag no emissions.

There discharge opening of the urine collection bag soft.

Bag storage capacity, represented by mL, specified by the manufacturer.

The amount is the amount of test when testing the reservoir bag strength and resistance to leakage, the amount of liquid injected into the storage bag (test volume) or applying a force (See 6.1).

4.1 Rated capacity

After the test according to 6.1, the nominal volume of the reservoir bag should not exceed the test volume.

4.2 No-load leakage

After pressing the test 6.2, the reservoir bag should not leak.

4.3 Loading leak

After pressing the test 6.3, the reservoir bag should not leak.

GB/T 17257.2-2009/ISO 8669-2:1996