

ICS 11.040.10

CCS C 45



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 18375.6-2004

**Prosthetics -- Structural testing of lower-limb prostheses -- Part
6: Loading parameters of supplementary structural tests**

假肢 下肢假肢的结构检验 第6部分：辅助结构检验中的加载参数

(English Translation)

Issued on: June 11, 2004

Implemented on: January 1, 2005

**Issued by: State Administration of Quality Supervision, Inspection and
Quarantine of the People Republic of China China National
Standardization Administration Committee**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies: lower limb prostheses static and dynamic testing process, in this process, addition to a special case, are to produce a composite with a test force test load. On the test sample depending on the load of each component of composite load Peak, these peaks occur during walking through the support phase when the different instantly.

Document data

International classification (ICS): 11.040.10.

Chinese classification (CCS): C45.

Estimated length of the Chinese original: 8,863 words.

Adopted international standard: ISO 10328-6-1996; IDT.

Standards quoted by this document

GB/T 14191-1993; GB/T 18375.3-2001