

ICS 75.080
CCS E 30



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 1884-2000

Replacing GB/T 1884-1992

**Crude petroleum and liquid petroleum products. Laboratory
determination of density. Hydrometer method [including
MODIFICATION 1]
(English Translation)**

Issued on: April 3, 2000

Implemented on: July 1, 2000

Issued by: State Quality and Technical Supervision

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies: usually measured in the laboratory using a glass petroleum liquid density meter (hereinafter referred to as density meter) crude oil, petroleum products and petroleum products and non-oil mixtures -20 ? density methods. These liquids Reid Vapor Pressure (RVP) of less than 100kPa. This standard applies to: Determination of transparent liquid flowing density, You can also use the right thermostatic bath, at above room temperature Determination of viscous liquid case, also for opaque liquids, reading the meniscus of liquid density meter dry pipe tangent at reading, and amend table 1.

Document data

International classification (ICS): 75.080.

Chinese classification (CCS): E30.

Estimated length of the Chinese original: 1,196 words.

Adopted international standard: ISO 3675-1998; MOD.

Supersedes the earlier edition GB/T 1884-1992.

Standards quoted by this document

GB/T 1885; GB/T 4756; SH/T 0316