

ICS 11.040.40

CCS C 45



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 19543-2004

**Prostheses -- Structural testing of hip units
(English Translation)**

Issued on: June 11, 2004

Implemented on: January 1, 2005

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine of the People Republic of China, China National
Standardization Administration Committee**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies: hip disarticulation prostheses hip and thigh components and assemblies test methods, This test method GB/T 18375 in lower limb prosthetics does not apply to other components. This standard specifies: right hip two different test surface (front, side) test force is applied to generate load components static strength test and dynamic strength tests. The components of the test sample the components of the load depending on the load peak, these peaks typically occur during walking with the support different moments.

Document data

International classification (ICS): 11.040.40.

Chinese classification (CCS): C45.

Estimated length of the Chinese original: 33,354 words.

Adopted international standard: ISO 15032-2000; IDT.

Standards quoted by this document

GB/T 14191-1993