

ICS 11.180
CCS C45



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 22752-2021

Replacing GB/T 22752-2008

Assistive product for person with disability - Grab bar

康复辅助器具 抓握杆

Issued on: April 30, 2021

Implemented on: April 30, 2021

**Issued by: State Administration for Market Regulation;
Standardization Administration of China**

Table of Contents

- 1 Scope
- 2 Normative references
- 3 Terms and definitions

Foreword

This standard was drafted in accordance with the rules given in GB/T 1.1-2009. This standard replaces GB/T 22752-2008 "Grip bar for assistive devices for the disabled". Compared with GB/T 22752-2008, the main technology The changes are as follows:

- Modified the scope of the standard (see Chapter 1, Chapter 1 of the.2008 edition);
- Modified the normative reference documents (see Chapter 2, 4.4, Chapter 2, 4.4 of the.2008 edition);
- Modified the definition of grip bar (see 3.1, 3.1 in.2008 edition);
- Modified the English corresponding words of the term (see 3.1, 3.2, 3.5, 3.1, 3.2, 3.5 in the.2008 edition);
- Deleted part of the requirements for materials (see 4.2.1, 4.2.2 of the.2008 edition);
- Modified the name and requirements of the strength of the connectors (see 4.3.3, 4.3.3 of the.2008 edition);
- Added fatigue strength requirements and test methods (see 4.3.4, 5.3.4);
- Modified the name of the installation (see 4.4,

1 Scope

GB/T 22752-2021 is the Chinese national standard on assistive product for person with disability - grab bar, in the field of health care technology. The /T suffix marks it as a recommended standard: it is not compulsory by itself, but it becomes binding as soon as a contract, a tender or a customer specification calls it up - which in practice is how most foreign buyers meet it. It was issued on 30 April 2021 by the State Administration for Market Regulation; Standardization Administration of China, and has been in force since 30 April

2021. Classification: ICS 11.180, CCS C45. This page is published from the official record of the standard held by the Chinese standards administration: the identification, the dates, the classification and the issuing body are taken from there. The clause text, the tables and the numeric limits are in the document itself, which is delivered complete in English translation.

2. It is usually a straight or curved rod.

3.2 Common grabbar It is installed on the floor or wall and has a gripping rod with two fixed support points. see picture 1.

a) The grip bar is installed on the floor

b) The grip bar is installed on the wall Figure

2 Normative references

The following documents are indispensable for the application of this document. For dated reference documents, only the dated version applies to this article. For undated reference documents, the latest version (including all amendments) is applicable to this document.

GB/T 191 Packaging, storage and transportation pictorial signs

GB/T 16886.1 Biological Evaluation of Medical Devices Part

3 Terms and definitions

The following terms and definitions apply to this document.

3.1 Grabbar A device that is fixed on the wall, wall and ground, or on the ground, wall and ceiling, and plays a supporting role when a person stands, moves or changes body position. Note