

ICS 13.160
CCS J 04



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 23716-2009

Human response to vibration -- Measuring instrumentation

人体对振动的响应 测量仪器

(English Translation)

Issued on: April 24, 2009

Implemented on: December 1, 2009

Issued by: Standardization Administration of China

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard provides for the evaluation of human response to vibration of the vibration measuring instrument performance specifications and tolerance limits. It includes the type evaluation, requires periodic testing and on-site inspections. And for on-site inspection of the vibration calibrator specifications. This standard specifies the vibration measuring instrument can be a single instrument, instrument combination or acquisition and analysis system based computer. This standard specifies the vibration measuring instruments can be applied to one or more of the following purposes vibration measurement: hand-transmitted vibration (see ISO 5349-1), whole-body vibration (see GB/T 13441. 1-2006, GB/T 13441. 2-2008, ISO 2631-4), at 0. 1Hz ~ 0. 5Hz frequency range, low-frequency whole-body vibration.

Document data

International classification (ICS): 13.160.

Chinese classification (CCS): J04.

Estimated length of the Chinese original: 80,888 words.

Adopted international standard: ISO 8041-2005, IDT.