

ICS 11.180.01

CCS C 45



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 24434-2025

Replacing GB/T 24434-2009

Commode chairs

坐便椅

Issued on: June 30, 2025

Implemented on: January 1, 2026

**Issued by: State Administration for Market Regulation;
Standardization Administration of the PRC**

Table of Contents

Preface	III
1 Scope	1
2 Normative references	1
3 Terms and Definitions	1
4 Categories	2
5 Technical Requirements	4
5.1 General requirements	4
5.2 Appearance requirements	4
5.3 Assembly requirements	4
5.4 Braking device	5
5.5 Stability	5
5.6 Strength requirements	5
5.7 Potty Requirements	5
6 Test conditions	6
6.1 General test conditions	6
6.2 Test equipment	6
7 Test Method	6
7.1 General requirements inspection	6
7.2 Appearance inspection requirements	6
7.3 Assembly requirements inspection	6
7.4 Braking device test	6
7.5 Stability test	7
7.6 Strength requirement test	7
7.7 Bedpan requirement test	9
8 Inspection Rules	9
8.1 Factory Inspection	9
8.2 Type inspection	9
8.3 Sampling and Decision Rules	10
9 Marking, packaging, transportation, storage 10.	10
9.1 Sign	10
9.2 Packaging	10
9.3 Transportation and storage	10

Foreword

This document is in accordance with the provisions of GB/T 1.1-2020 "Guidelines for standardization work Part 1: Structure and drafting rules for standardization documents" Drafting.

This document replaces GB/T 24434-2009 "Commode Chair (Stool)". Compared with GB/T 24434-2009, except for structural adjustments and editorial In addition to the modifications, the main technical changes are as follows.

- Changed the definition of commode chair (see 3.1, 3.1 of the 2009 edition);
- Added the definition of maximum user quality (see 3.12);
- Changed the classification and deleted the model name (see Chapter 4, Chapter 4 of the 2009 edition);
- Added items in the assembly requirements (see 5.3.6, 5.3.7);
- Changed the test conditions (see Chapter 6, 6.1 of the 2009 edition);
- Changed the load-bearing mass in the stability test (see 7.5, 6.4 of the 2009 edition);
- Changed some static load test parameters and added static load test items (see 7.6.1, 6.3 of the 2009 edition);
- Added backrest impact resistance test (see 7.6.2.2);
- Increased durability requirements and test methods (see 5.6.3, 7.6.3);
- Changed the water temperature setting for the bedpan temperature change test (see 7.7.2.2, 6.7.2 of the 2009 edition).

Please note that some of the contents of this document may involve patents. The issuing organization of this document does not assume the responsibility for identifying patents.

This document is issued by the Ministry of Civil Affairs of the People's Republic of China.

This document is under the jurisdiction of the National Technical Committee for Standardization of Rehabilitation and Special Equipment for Persons with Disabilities (SAC/TC148).

This document was drafted by: China Center for Assistive Devices and Equipment for Persons with Disabilities, Foshan Kaiyang Medical Equipment Co., Ltd., Henan Xiangyu Medical Equipment Co., Ltd.

Co., Ltd., Foshan Dongfang Medical Equipment Factory Co., Ltd., Shaanxi Medical Device Quality Inspection Institute, Jiangyin Xinsheng Medical Equipment Co., Ltd.

Co., Ltd., China National Testing Holding Group Shaanxi Co., Ltd., and Hengshui New

Century Medical Technology Co., Ltd.

The main drafters of this document are: Zhang Hongtao, Zhang Jian, Li Yu, Wang Hao, Zhou Mi, Liao Youfeng, He Chuan, Wu Ruiping, Xu Wenfeng, Huang Bing, Sun Jianwu, Gu Deyin, Min Nan, and Zhu Hui.

The previous versions of this document and the documents it replaces are as follows.

- First published in 2009 as GB/T 24434-2009;
- This is the first revision. Commode chair

1 Scope

This document defines the terms and definitions of toilet chairs, and stipulates the classification, technical requirements, inspection rules, marking, packaging, transportation, Storage, describes the corresponding test method.

This document applies to fixed toilet chairs and folding toilet chairs with a user mass of not less than 35kg and not more than 150kg as specified by the manufacturer.

Design, production and testing of commode chairs, commode chairs with wheels, simple commode chairs and other commode chairs.

2 Normative references

GB/T 191

GB/T 16432

GB/T 43386

3 Terms and Definitions

The terms and definitions defined in GB/T 16432 and the following apply to this document.

3.1 commode chair

Chair with collection container for both bowel and bladder, with or without castors.

Note. This includes hygienic chairs, commode wheelchairs and commode chairs for showers.

3.2 Fixed commode chair

Commode chairs (3.1) that are not foldable and consist of armrests, seat, pan and frame.

3.3 folding commode chair

A foldable commode chair (3.1) consisting of armrests, seat, pan and frame.

3.4

A movable commode chair (3.1) consisting of a backrest, armrests, seat, bedpan, frame and castors.

3.5 Simple commode chair simplecommodechair

Commode chair (3.1) consisting of seat, pan and frame, also called commode stool.