

ICS 19.100
CCS J 04



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 25758.3-2010

**Non-destructive testing -- Characteristics of focal spots in
industrial X-ray systems for use in non-destructive testing --
Part 3: Slit camera radiographic method**

无损检测 工业X射线系统焦点特性 第3部分：狭缝照相机射线照相方法
(English Translation)

Issued on: December 23, 2010

Implemented on: October 1, 2011

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine of the People's Republic of China, Standardization
Administration of the People's Republic of China**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the use of a slit camera radiographic method. Measuring tube voltage up to 500 kV industrial X-ray system is greater than 0.1 mm focal spot size method. The measurement voltage is limited by section 200 kV.

Document data

International classification (ICS): 19.100.

Chinese classification (CCS): J04.

Estimated length of the Chinese original: 10,131 words.

Adopted international standard: EN 12543-3-1999, IDT.

Standards quoted by this document

GB/T 19348.1; GB/T 25758.1