

ICS 17.220.20
CCS N 25



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 3927-2008

Replacing GB/T 3927-1983

Direct-current potentiometers

直流电位差计

(English Translation)

Issued on: August 6, 2008

Implemented on: March 1, 2009

Issued by: Standardization Administration of China

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard applies to the level of accuracy of 0.0001 ~ 0.1 [1ppm ~ 1000ppm (parts per million), 1×10^{-6} ~ 1×10^{-3} (scientific notation, scientific notation)] of the DC potentiometer, and applies to the potential difference meter containing auxiliary equipment. This standard does not apply to automatic zero balancing potentiometer, do not apply to those obtained by the balanced detector portion of the scale indicated value potentiometer, do not apply potentiometer used in conjunction with an external auxiliary equipment.

Document data

International classification (ICS): 17.220.20.

Chinese classification (CCS): N25.

Estimated length of the Chinese original: 17,112 words.

Adopted international standard: IEC 60523-1997, IDT.

Supersedes the earlier edition GB/T 3927-1983.