

ICS 17.06
CCS N 61



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 4168-1992

Replacing GB 4168-1984

**Non-automatic balance. Beam balance
(English Translation)**

Issued on: November 4, 1992

Implemented on: August 1, 1993

Issued by: State Bureau of Technical Supervision

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the non-automatic balance in leveraged balance (hereinafter referred to as the balance) of the basic parameters, technical requirements, test methods, inspection rules and signs, packaging, transportation, storage. This standard applies to the accuracy level of $1 \times 10^{-7} \sim 1 \times 10^{-4}$, the largest weighing in the range of 1 g ~ 500 kg blade-supporting single-leveraged balance. This standard does not apply to the vacuum balance, heat balance, the balance remote control, automatic recording balance and the loading balance and according to the agreement made in the export balance.

Document data

International classification (ICS): 17.06.

Chinese classification (CCS): N61.

Estimated length of the Chinese original: 8,831 words.

Supersedes the earlier edition GB 4168-1984.

Standards quoted by this document

ZBY 002; ZBY 032; ZB N61 009; ZB N61 010; ZB N61 011; ZB N61 013; JJG 98