

ICS 11.180.01

CCS C 45



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 45933-2025

**Basic configuration of assistive products for the senior care
organization**

养老机构康复辅助器具基本配置

Issued on: June 30, 2025

Implemented on: October 1, 2025

**Issued by: State Administration for Market Regulation;
Standardization Administration of the PRC**

Table of Contents

Preface	III
1 Scope	1
2 Normative references	1
3 Terms and Definitions	1
4 Basic Requirements	1
5 Requirements	2
5.1 Public Activity Area	2
5.2 Room	2
5.3 Toilet and bathroom	2
5.4 Rehabilitation training room (indoor/outdoor)	3
5.5 Infirmary/Nursing Station	3
5.6 Dining Space	3
5.7 Reading Room	3
5.8 Entertainment Room	3
Appendix A (Informative) Basic Configuration List of Personal Rehabilitation Assistive Devices	4
Appendix B (Normative) Basic Configuration List of Rehabilitation Assistive Devices in Each Area of Nursing Home	5
References	6

Foreword

This document is in accordance with the provisions of GB/T 1.1-2020 "Guidelines for standardization work Part 1: Structure and drafting rules for standardization documents" Drafting.

Please note that some of the contents of this document may involve patents. The issuing organization of this document does not assume the responsibility for identifying patents.

This document is issued by the Ministry of Civil Affairs of the People's Republic of China.

This document is under the jurisdiction of the National Technical Committee for Standardization of Rehabilitation and Special Equipment for Persons with Disabilities (SAC/TC148).

This document was drafted by: Hubei Rehabilitation Assistive Devices Technology Center, China Rehabilitation Assistive Devices Association, Hubei Institute of Standardization and

Quality, Hubei Health Care Group Co., Ltd., Xiamen Jinyitong Precision Manufacturing Co., Ltd., Hubei Provincial Department of Civil Affairs, Hebei Weizhuo Medical Care Technology Co., Ltd.

Co., Ltd., Wuhan Biying Biotechnology Co., Ltd., Jiangsu Disabled Persons Rehabilitation Center, Sichuan Rehabilitation Assistive Devices Technology Service Center, Hangzhou Huier Hearing Technology Equipment Co., Ltd., Shenzhen Huashengyuan Urban Service Group Co., Ltd., and Wuhan Xinmeiyu Technology Co., Ltd.

The main drafters of this document are: Shuang Kang, Hui Jie, Yan Heping, Zeng Donglin, Mao Yong, Zhang Pengcheng, Zhang Zhengfei, Xiong Chaoyang, Zhao Zhongtao, Xiong Wei, Xiong Wen, Liu Yuling, Liu Rui, Luo Laishan, Wang Wenjun, Xu Yanhua, Ding Fan, Ren Yufeng, Hu Jiao, Liu Wan, Jia Lianwen, Ma Linghua, Huang Jiatao, Liu Hang, Wei Kaijun, Wang Lingna, Feng Qiang, Ren Dingwei, Cai Jingxuan, Zhu Yan, He Miaomiao, Liu Yujiao, Zhang Liu, Tu Lei, Chen Shucong, Li Shuang, Liu Wei, and Wu Tao.

Basic configuration of rehabilitation assistive devices in nursing homes

1 Scope

This document specifies the basic requirements and configuration requirements for rehabilitation assistive devices in nursing homes, and provides the requirements for individual use and various nursing home use.

Basic configuration list of rehabilitation assistive devices for regional space.

This document applies to the basic configuration of rehabilitation assistive devices in nursing homes.

2 Normative references

GB/T 42195

GB 50763

3 Terms and Definitions

The following terms and definitions apply to this document.

3.1 assistive devices

Rehabilitation assistive devices Products and technologies that improve human performance and alleviate functional impairments.

Note 1. Rehabilitation assistive devices include instruments, equipment, devices and software.

Note 2 to entry. Rehabilitation assistive devices are either custom-made or ready-made.

[Source. GB/T 16432-2025, 3.3, modified]

4 Basic Requirements

4.1 Elderly care institutions should be oriented towards meeting the needs of the elderly, serve the elderly, be suitable for the abilities of the elderly, and provide corresponding rehabilitation support. Assistive devices.

4.2 The configuration of rehabilitation assistive devices in nursing homes should be based on the principle of creating a safe and barrier-free environment to facilitate the activities and use of the elderly.

4.3 Nursing homes should assess the abilities of the elderly before they move in, determine their ability levels, and recommend Individuals can configure their own rehabilitation assistive devices for personal use. The configuration catalog is shown in Appendix A.

4.4 Nursing homes should be divided into different areas and equipped with relevant rehabilitation assistive devices according to Appendix B.

4.5 The design and installation of barrier-free buildings and facilities should comply with the requirements of GB 50763.

4.6 Nursing homes may apply intelligent rehabilitation assistive device technology.