



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 6730.63-2024

**Iron ores - Determination of aluminum, calcium, magnesium,
manganese, phosphorus, silicon and titanium content - Inductively
coupled plasma atomic emission spectrometric method**

Issued on: May 28, 2024

Implemented on: December 1, 2024

Issued by: SAMR; SAC

Table of Contents

Foreword	3
Introduction	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	8
4 Principle	8
5 Reagents and materials	8
6 Instruments and equipment	11
7 Sampling and specimen preparation	12
8 Analytical procedures	12
9 Calculation of results	15
10 Test report	21
Appendix A (Informative) Components of GB/T 6730	22
Appendix B (Informative) Comparison of structure numbers of this document with ISO 11535.2006	27
Appendix C (Informative) Recommended solution concentrations for calibration curves ..	28
Appendix D (Normative) Performance test of plasma spectrometer	29
Appendix E (Informative) Derivation of repeatability and allowable deviation equations ...	32
Appendix F (Normative) Acceptance procedure for specimen analysis result	33

1 Scope

This document specifies the determination of aluminum, calcium, magnesium, manganese, phosphorus, silicon, titanium content by inductively coupled plasma emission spectrometry.

2 Normative references

The contents of the following documents constitute the essential terms of this document through normative references in the text. Among them, for dated references, only the version corresponding to that date applies to this document; for undated references, the

5.5 Hydrochloric acid, 1 + 1.

The calibration solution is the solution required for drawing the calibration curve of the analyzed elements. Its concentration range in the solution is expressed in micrograms per milliliter, which depends on the performance parameters and linear sensitivity of the

instrument. At least 10 solutions are required to cover the measurement range (mass fraction) shown in Table 1.

8.4 Determination The ICP spectrometer shall first be initially adjusted, according to the manufacturer's recommendations and laboratory quantitative analysis operations.

Try to optimize the instrument parameters in each round as much as possible.

9.3 Standardization of calibration curve (drift correction)

The regular inspection and correction of the calibration curve shall be carried out as follows.

Take two calibration solutions, namely the one with the lowest and highest analyte content.

In the drawn calibration curve, the intensities of the two calibration solutions are measured; the correction coefficients α and β are calculated according to formulas (4)

and (5).