

ICS 17.140.10
CCS A 42



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/T 7583-1987

**Acoustics. Pure tone air conduction threshold audiometry. For
hearing conservation purposes
(English Translation)**

Issued on: March 27, 1987

Implemented on: November 1, 1987

Issued by: National Bureau of Standards

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the staff to monitor the hearing in a noisy environment for the purpose of pure tone air conduction hearing threshold necessary conditions and test procedures determined, including the audiometer calibration of audiometric indoor maximum permissible ambient noise level and ready to work. This standard specifies the hand control and automatic recording audiometer audiometer hearing thresholds measured at a fixed frequency method, computer or other automated equipment can also be used to listen to a threshold determination. This standard specifies the test method used to listen to records obtained to determine the hearing threshold level and to make the provisions of the audiogram.

Document data

International classification (ICS): 17.140.10.

Chinese classification (CCS): A42.

Estimated length of the Chinese original: 9,992 words.

Adopted international standard: ISO 6189-1983, NEQ.