

ICS 11.020
CCS C04



**NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GB/Z 40666-2021

**Specification for operation technologies of traditional Chinese
medicineDermatologyVapor bath of Chinese materia medica**

中医技术操作规范 皮肤科 中药蒸气浴

Issued on: October 11, 2021

Implemented on: October 11, 2021

**Issued by: State Administration for Market Regulation;
Standardization Administration of China**

Table of Contents

- 1 Scope
- 2 Normative references
- 3 Terms and definitions
- 4 Preparation before operation
- 5 Operation steps

Foreword

This document is in accordance with GB/T 1.1-2020 "Guidelines for Standardization Part 1. Structure and Drafting Rules of Standardization Documents". To be drafted. Please note that some of the contents of this document may involve patents. The issuing agency of this document is not responsible for identifying patents. This document was proposed by the State Administration of Traditional Chinese Medicine. This document is under the jurisdiction of the National Standardization Technical Committee of Traditional Chinese Medicine (SAC/TC478). This document was drafted by: Wuhan Hospital of Integrated Traditional Chinese and Western Medicine, Chinese Association of Chinese Medicine, Second Affiliated Hospital of Hunan University of Traditional Chinese Medicine, Jiangxi Chinese Medicine The Second Affiliated Hospital of Medical University, Shenyang Hospital of Integrated Traditional Chinese and Western Medicine, Beijing Hospital of Traditional Chinese Medicine Affiliated to Capital Medical University, Guangdong Provincial Hospital of Traditional Chinese Medicine. The main drafters of this document. Duan Yiqun, Li Kai, Zhou Xiaoyong, Su Xiangfei, Yang Zhibo, Liu Qiao, Fan Ruiqiang, Chen Qingyan, Zhou Dongmei. GB/T 40666-2021 Traditional Chinese Medicine Technical Code of Practice

1 Scope

GB/Z 40666-2021 is the Chinese national standard on specification for operation technologies of traditional chinese medicinedermatologyvapor bath of chinese materia medica, in the field of health care technology. The /Z suffix marks it as a guiding technical document: it does not prescribe requirements that can be certified against, but sets out the technique, the method or the state of the art that the standards bodies recommend following. It was issued on 11 October 2021 by the State Administration for Market Regulation; Standardization Administration of China. As a guiding technical document it carries no separate date of entry into force: it applies from publication. Classification: ICS 11.020, CCS C04. This page is published from the official record of the standard held by the Chinese standards administration: the identification, the dates, the classification and the

issuing body are taken from there. The clause text, the tables and the numeric limits are in the document itself, which is delivered complete in English translation.

This document stipulates the preparation, operation steps, precautions and contraindications before the operation of the traditional Chinese medicine steam bath. This document is applicable to the technical operation of traditional Chinese medicine steam bath.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

The following terms and definitions apply to this document.

3.1 Chinese medicine steam bath vapor bath of Chinese materia medica Under the guidance of traditional Chinese medicine theory, combined with modern traditional Chinese medicine steam bath therapy instrument, the traditional Chinese medicine decoction is heated and converted into traditional Chinese medicine vapor, and the patient is fumigated by the machine. A traditional Chinese medicine external treatment method that achieves the purpose of treatment.

4 Preparation before operation

4.1 Therapeutic equipment and supplies Therapeutic equipment and supplies include. traditional Chinese medicine steam bath therapy instrument, traditional Chinese medicine liquid, disposable steam bath clothes, gaskets, trichloro effervescent disinfection tablets, Measuring cup.

4.2 Preparation of traditional Chinese medicine liquid The decoction of Chinese medicinal materials or crude powder is decocted in water, and the slag is removed and the juice is taken for later use.

4.3 Disinfection and cleaning of the cabin of the traditional Chinese medicine steam bath therapy instrument Before each use, use 1500mg of trichloro effervescent disinfectant tablets, add 2000mL of water, and dilute to a concentration of 750mg/L liquid. In the poison inlet, add 2000mL water in a measuring cup into the clean inlet. Disinfect and clean after staying for 5 minutes.

4.4 Environment Keep room temperature about 25°C~28°C, relative humidity 60%~70%. Maintain a good ventilation environment.

5 Operation steps

5.1 Warm up the traditional Chinese medicine steam bath therapeutic apparatus within 30 minutes before treating the patient.

5.2 Check the patient's name, age, disease diagnosis, treatment order and other information.

5.3 Turn on the power and check whether the Chinese medicine steam bath treatment instrument is operating normally. GB/T 40666-2021

chinastandards.org - PREVIEW