

ICS 13.1
CCS C 57



**NATIONAL OCCUPATIONAL HEALTH STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

GBZ/T 201.5-2015

**Radiation shielding requirements for radiotherapy rooms -- Part
5: Radiotherapy room of proton accelerators
(English Translation)**

Issued on: November 16, 2015

Implemented on: May 1, 2016

Issued by: National Health and Family Planning Commission

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This Standard specifies the proton accelerator for radiotherapy treatment rooms radiation shielding requirements. This Standard applies to 70 MeV ~ 250 MeV proton accelerator for radiotherapy treatment rooms. This Standard does not apply to scientific research by proton accelerator room, not suitable for the production of radionuclides proton accelerator room.

Document data

International classification (ICS): 13.1.

Chinese classification (CCS): C57.

Estimated length of the Chinese original: 16,165 words.

Standards quoted by this document

GBZ/T 201.1-2007; GBZ/T 201.2