

HJ

INDUSTRY STANDARD OF THE PEOPLE'S REPUBLIC OF CHINA

HJ 1323-2023

**Determination of lead-210 in water - Resin separation- β counter
method**

水中铅-210 的分析方法 冠醚树脂分离- β 计数器法
(English Translation)

Issued on: December 5, 2023

Implemented on: January 1, 2024

Issued by: Ministry of Ecology and Environment

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the crown ether resin separation-? counter method for the determination of lead-210 in water. This standard applies to the analysis and measurement of lead-210 in surface water, groundwater, drinking water, seawater and wastewater. Under typical conditions, the detection limit of the method can reach 2.0 mBq/L. See Appendix A.1 for specific content.

Document data

Estimated length of the Chinese original: 14,129 words.