

HJ

INDUSTRY STANDARD OF THE PEOPLE'S REPUBLIC OF CHINA

HJ 1384-2024

**Technical specification for continuous monitoring of fixed NaI
(TI) gamma spectrometer**

固定式碘化钠 γ 谱仪连续监测技术规范
(English Translation)

Issued on: December 20, 2024

Implemented on: February 1, 2025

Issued by: Ministry of Ecology and Environment

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the measurement system, ? radiation air absorption dose rate, ? spectrum, data statistics and recording, quality assurance and other aspects of the fixed sodium iodide ? spectrometer of the radiation environment air automatic monitoring station. This standard applies to the fixed sodium iodide ? spectrometer of the air automatic monitoring station for environmental quality monitoring and nuclear facility surrounding environment monitoring in the field of radiation monitoring. Fixed sodium iodide ? spectrometers or drop-type sodium iodide ? spectrometers equipped for other purposes can be implemented as a reference.

Document data

Estimated length of the Chinese original: 11,112 words.