

HJ

INDUSTRY STANDARD OF THE PEOPLE'S REPUBLIC OF CHINA

HJ 915.3-2024

**Technical specifications for operation and maintenance of
surface water quality automated monitoring station
(Conventional five parameters, CODMn, NH₃-N, TP, TN)**

地表水水质自动监测站（常规五参数、CODMn、NH₃-N、TP、TN）运行维护技术规范

(English Translation)

Issued on: December 25, 2024

Implemented on: July 1, 2025

Issued by: Ministry of Ecology and Environment

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the technical requirements for inspection and maintenance, quality control, abnormal situation handling and operation records of automatic surface water quality monitoring stations (hereinafter referred to as water stations). This standard applies to the inspection, maintenance, quality assurance and quality control of water stations. It is applicable to automatic water quality monitoring instruments such as the conventional five parameters (water temperature, pH, dissolved oxygen, conductivity, turbidity), permanganate index (CODMn), ammonia nitrogen (NH₃?N), total phosphorus (TP), and total nitrogen (TN). Automatic water quality monitoring instruments for other monitoring projects can refer to this standard for operation and maintenance.

Document data

Estimated length of the Chinese original: 17,186 words.