

ICS 13.060
CCS Z 50

HJ

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

HJ/T 345-2007

**Water quality. Determination of iron-phenanthroline
spectrophotometry**

水质 铁的测定 邻菲罗啉分光光度法

(English Translation)

Issued on: March 10, 2007

Implemented on: May 1, 2007

Issued by: Ministry of Ecology and Environment

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard applies to surface water, groundwater and wastewater in the determination of iron. Methods minimum detectable concentration of 0.03mg/L, detection limit of 0.12mg/L, measured ceiling 5.00mg/L. Of iron ions is greater than 5.00mg/L in water samples, may be appropriate after dilution were determined according to the method.

Document data

International classification (ICS): 13.060.

Chinese classification (CCS): Z50.

Estimated length of the Chinese original: 8,810 words.