

ICS 17.220
CCS A 56

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1127-2004

Replacing JJG 127-1986

**Calibration Specification for RF impedance/Meterial Analyzers
(English Translation)**

Issued on: September 21, 2004

Implemented on: March 21, 2005

**Issued by: The General Administration of Quality Supervision, Inspection
and Quarantine issued, This specification was approved by the General
Administration of Quality Supervision, Inspection and Quarantine on
September 21,.2004, and**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to the calibration frequency range 1MHz ~ 3.0GHz, the basic working principle of the RF voltage and current ratio are given for various models of high frequency impedance and RF impedance/material analyzer calibration, the principles also apply to the same kind of LCR table.

Document data

International classification (ICS): 17.220.

Chinese classification (CCS): A56.

Estimated length of the Chinese original: 1,514 words.

Supersedes the earlier edition JJG 127-1986.

Standards quoted by this document

JJG 2011-1987; JJF 1059-1999