

ICS 17.200
CCS A 54

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1178-2007

**Calibration Specification of Fixed-Point Devices for Standard
Platinum Resistance Thermometer
(English Translation)**

Issued on: June 14, 2007

Implemented on: September 14, 2007

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to -189. 3442°C ~ 961. 78°C temperature range for indexing standard platinum resistance thermometers and platinum resistance thermometer calibration of fixed-point devices.

Document data

International classification (ICS): 17.200.

Chinese classification (CCS): A54.

Estimated length of the Chinese original: 18,174 words.

Standards quoted by this document

JJG 160-1992; JJG 716-1991; JJG 859-1994; JJG 985-2004