

ICS 17.220
CCS A 55

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1217-2009

**Calibrator Specification for Electrosurgical Generator
(English Translation)**

Issued on: May 18, 2009

Implemented on: August 18, 2009

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to newly manufactured, use and repair, the work is quite a range 0.3MHz ~ 5.0MHz rate unipolar, bipolar electrosurgical medical calibration is quite high. This specification does not apply to a unipolar mode the maximum power is less than 50W of John the calibration of medical high-frequency electric knife, does not apply to dental bipolar mode, gynecology and dermatology, and other special high-frequency electric knife calibration.

Document data

International classification (ICS): 17.220.

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 20,246 words.

Standards quoted by this document

JJF 1001-1998; JJF 1071-2000; JJF 1059-1999; GB 9706.1-2007; GB 9706.4-1999