

ICS 17.18
CCS A 60

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1290-2011

**Calibration specification for Particulate Analyzer
(English Translation)**

Issued on: June 14, 2011

Implemented on: September 14, 2011

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to the use of insoluble particles assay (resist method), measuring particle size in the range of 2µm ~ 100 µm, particle concentration range for each 1mL containing 10 to 100 000 particles of particle detector calibration.

Document data

International classification (ICS): 17.18.

Chinese classification (CCS): A60.

Estimated length of the Chinese original: 14,160 words.

Standards quoted by this document

JJF 1001-1998; JJF 1094-2002; PRC Pharmacopoeia (2010)