

ICS 17.020
CCS A 61

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1316-2011

**Calibration Specification for Blood Viscometers
(English Translation)**

Issued on: September 20, 2011

Implemented on: December 20, 2011

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to a constant shear rate, the measurement range of the shear rate 1 s-1 ~ 200 s-1, the shear stress 10mPa ~ 1000mPa, measuring dynamic viscosity of whole blood and plasma blood viscometer (hereinafter referred to as the instrument) of calibration.

Document data

International classification (ICS): 17.020.

Chinese classification (CCS): A61.

Estimated length of the Chinese original: 14,116 words.

Standards quoted by this document

GB/T 10247-2008