

ICS 17.180
CCS A 60

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1329-2011

**Calibration Specification for Instantaneous Spectral Instruments
(English Translation)**

Issued on: December 28, 2011

Implemented on: March 28, 2012

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to the wavelength range of 250nm ~ 1750nm, the source of the measured flash duration ?? 1×10^{-4} transient spectrometer calibration. Other types of transient spectrometer can be calibrated with reference to the specification.

Document data

International classification (ICS): 17.180.

Chinese classification (CCS): A60.

Estimated length of the Chinese original: 22,267 words.

Standards quoted by this document

JJG 383-2002; JJG 768-2005; JJF 1032-2005; CIE 105-1993