

ICS 17.040
CCS A 52

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1331-2011

Replacing JJG 396-2002

**Calibration Specification for Inductive Micrometers
(English Translation)**

Issued on: December 28, 2011

Implemented on: June 28, 2012

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to resolution (scale value) (0.01 ~ 10) μm , nominal range (-1 000 ~ +1000) μm inductance micrometer calibration.

Document data

International classification (ICS): 17.040.

Chinese classification (CCS): A52.

Estimated length of the Chinese original: 23,223 words.

Supersedes the earlier edition JJG 396-2002.

Standards quoted by this document

GB/T 26094-2010; GB/T 26095-2010; GB/T 26097-2010