

CCS A 55

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1517-2015

**Calibration Specification for Contactless Electrostatic Voltage.
Measuring Instruments
(English Translation)**

Issued on: April 10, 2015

Implemented on: July 10, 2015

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This Standard applies to the measurement of the surface voltage not exceeding 20kV electrostatic voltage non-contact measuring instrument (also known as non-contact electrostatic voltmeter, hereinafter referred to as static tables) calibration. This specification does not apply to contact electrostatic voltage meter calibration.

Document data

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 21,287 words.