

ICS 17.200
CCS A 54

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 1637-2017

Replacing JJG 351-1996

**Calibration Specification for Base Metal Thermocouples
(English Translation)**

Issued on: September 26, 2017

Implemented on: March 26, 2018

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to measuring range $-40^{\circ}\text{C} \sim 1200^{\circ}\text{C}$, length not less than 500mm, removable nickel-chromium-nickel-silicon (K type), nickel-chromium-silicon-nickel-silicon-magnesium (N type), nickel-chromium-copper-nickel (E Type), iron - copper nickel (J type) cheap metal thermocouple (hereinafter referred to as school thermocouple) calibration.

Document data

International classification (ICS): 17.200.

Chinese classification (CCS): A54.

Estimated length of the Chinese original: 27,242 words.

Supersedes the earlier edition JJG 351-1996.

Standards quoted by this document

JJF 1007-2007; GB/T 4989-2013; GB/T 16701-2010; GB/T 16839.1; GB/T 16839.2; GB/T 30429-2013