

JJF

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJF 2119-2024

**Calibration Specification for Atmospheric pressure test
chambers
(English Translation)**

Issued on: June 14, 2024

Implemented on: December 14, 2024

Issued by: State Administration for Market Regulation

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard applies to low-pressure test chambers with an absolute pressure range of 0.1 kPa~120 kPa and an absolute pressure range of 0.1 kPa~120 kPa and a temperature range of -80°C~200 °C range of high and low temperature low pressure test chamber (hereinafter referred to as "test chamber") calibration.

Document data

Estimated length of the Chinese original: 23,277 words.