

ICS 17.220
CCS A 55

JJG

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJG 1049-2009

**Verification Regulation of Alternating Tesla-Meter for Weak
Magnetic Field
(English Translation)**

Issued on: July 10, 2009

Implemented on: October 10, 2009

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to the electromagnetic induction magnetometer, alternating Hall magnetometer magnetic field range 1PT ~ 1MT, Frequency Range 10HZ ~ 10KHZ weak alternating magnetic field strength meter for initial verification, testing and use of follow-up inspection.

Document data

International classification (ICS): 17.220.

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 13,162 words.

Standards quoted by this document

JJF 1013-1989