

ICS 17.240
CCS A 58

JJG

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJG 1053-2009

**Verification Regulation of Radiation Source used in
60KV~300KV X-ray Radiotherapy
(English Translation)**

Issued on: October 9, 2009

Implemented on: January 9, 2010

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to 60kV ~ 300kV X-ray radiation therapy of initial verification, testing and use of follow-up inspection. Radiotherapy dose rate radiation output range is generally $(10^{-2} \sim 10)$ Gy/min.

Document data

International classification (ICS): 17.240.

Chinese classification (CCS): A58.

Estimated length of the Chinese original: 12,151 words.

Standards quoted by this document

JJF 1001-1998; JJG 912-1996; GB/T 17857-1999; " the absorption of photons and electron beam dosimetry International Practical Provisions" [IAEA Technical Report No. 227 (1998 Second Edition)]