

ICS 17.220
CCS A 55

JJG

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJG 124-2005

Replacing JJG 124-1993

**Amperemeter, Voltmeters, Wattmeters and Ohmmeters
(English Translation)**

Issued on: October 9, 2005

Implemented on: April 9, 2006

**Issued by: The General Administration of Quality Supervision, Inspection
and Quarantine issued, This regulation was approved by the General
Administration of Quality Supervision, Inspection and Quarantine on
October 9, 2005, and**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to Direct acting indicating analogue DC and AC (frequency 40Hz ~ 10kHz) ammeter, voltmeter, power meter, and an ohmmeter (resistance 1?? ~ 1M??) and measuring current, voltage and resistance of the multimeter (hereinafter referred to as instrumentation) for the first time test, follow-up testing and use of the test. This procedure does not apply to automatic recording type instruments, digital instruments, electronic instruments, the average voltage meter, peak voltage meter, leakage current meter, three-phase power meter and the voltage is higher than 600V electrostatic voltmeter test.

Document data

International classification (ICS): 17.220.

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 27,256 words.

Supersedes the earlier edition JJG 124-1993.

Standards quoted by this document

JJF 1001; JJF 1059; GB/T 7676.1; GB/T 7676.2; GB/T 7676.3; GB/T 7676.4; GB/T 7676.5; GB/T 7676.6; GB/T 7676.7; GB/T 7676.8; GB/T 7676.9