

ICS 17.220  
CCS A 55

**JJG**

**INDUSTRY STANDARD OF THE  
PEOPLE'S REPUBLIC OF CHINA**

**JJG 125-2004**

Replacing JJG 125-1986

---

**Verification Regulation of D.C.Bridges  
(English Translation)**

**Issued on: September 21, 2004**

**Implemented on: March 21, 2005**

---

**Issued by: The General Administration of Quality Supervision, Inspection  
and Quarantine issued, This regulation was approved by the General  
Administration of Quality Supervision, Inspection and Quarantine on  
September 21, 2004, and**

## Table of Contents

|                                         |   |
|-----------------------------------------|---|
| What this standard covers .....         | 1 |
| Document data .....                     | 2 |
| Standards quoted by this document ..... | 3 |

### What this standard covers

This standard applies to the resistance measurement limit of less than  $10^8 \Omega$ , the level of accuracy equal to or less than 0.005 resistive type DC bridge (hereinafter referred to as the bridge) initial verification, testing and use of the follow-up inspection. This procedure does not apply to automatic test bridge, semi-bridge, current comparator bridge, digital bridge and other special-purpose bridge.

### Document data

International classification (ICS): 17.220.

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 33,314 words.

Supersedes the earlier edition JJG 125-1986.

### Standards quoted by this document

JJF 1059-1999; JJF 1015-2002; GB 4793-1995; GB 3930-1983; IEC 60564-1997