

ICS 17.180
CCS A 60

JJG

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJG 247-2008

Replacing JJG 247-1991

**Verification Regulation of Standard Incandescent Lamp for Total
Luminous Flux
(English Translation)**

Issued on: March 25, 2008

Implemented on: September 25, 2008

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to initial verification AC bridge, follow-up test and use the test. AC bridge operating frequency is limited to within the range of 20Hz to 100kHz, the working voltage is less than 500V, accuracy range of 0.001 and below. This procedure does not apply to high-voltage bridge, verification LCR impedance meter and automatic balancing of the bridge.

Document data

International classification (ICS): 17.180.

Chinese classification (CCS): A60.

Estimated length of the Chinese original: 23,228 words.

Supersedes the earlier edition JJG 247-1991.

Standards quoted by this document

GB 15039-1994; GB 15040-1994; JJG 213-2003; JJF 1002-1998; JJF 1059-1999