

ICS 17.220
CCS A 55

JJG

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJG 441-2008

Replacing JJG 441-1986

**Verification Regulation of Alternating Current Bridge
(English Translation)**

Issued on: April 28, 2008

Implemented on: October 23, 2008

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to initial verification AC bridge, follow-up test and use the test. AC bridge operating frequency is limited to within the range of 20Hz to 100kHz, the working voltage is less than 500V, accuracy range of 0.001 and below. This procedure does not apply to high-voltage bridge, verification LCR impedance meter and automatic balancing of the bridge.

Document data

International classification (ICS): 17.220.

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 25,230 words.

Supersedes the earlier edition JJG 441-1986.

Standards quoted by this document

JJF 1001-1998; JJG 2073-1990; JJG 2075-1990; JJG 2076-1990; JJG 183-1992; JJG 726-1991; IEC 477.2-1979