

ICS 17.220
CCS A 55

JJG

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJG 495-2006

Replacing JJG 495-1987

**Verification Regulation of DC Magnetoelectric Galvanometers
(English Translation)**

Issued on: December 8, 2006

Implemented on: June 8, 2007

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to the DC magnetic lines galvanometer verification procedures for DC magnetic lines of the galvanometer (hereinafter referred to as the galvanometer) of the initial verification, testing and use of follow-up inspection. This procedure does not apply to: a) the measurement process does not directly by the position of the movable part of galvanometer readings, such as optical zoom galvanometer, electronic amplification galvanometer and optical recorder galvanometer, b) attached measuring device dedicated galvanometer, c) energy constant is less than 1×10^{-16} Ws/mm galvanometer, d) oscillation periods less than 0.5s galvanometer.

Document data

International classification (ICS): 17.220.

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 23,276 words.

Supersedes the earlier edition JJG 495-1987.

Standards quoted by this document

AQSIQ Announcement No. 187 of 2006