

INDUSTRY STANDARD OF THE PEOPLE'S REPUBLIC OF CHINA

JJG 531-2003

Replacing JJG 531-1988

Verification Regulation of The DC Resistive Volt Ratio Box (English Translation)

Issued on: May 12, 2003

Implemented on: November 12, 2003

Issued by: The General Administration of Quality Supervision, Inspection and Quarantine issued, This regulation was approved by the General Administration of Quality Supervision, Inspection and Quarantine on May 12, 2003, and

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard applies to a fixed rate, the upper limit of the rated input voltage is less than or equal to 1. 5kV, the level of accuracy of 0. 0005 ~ 0. 1 DC resistor divider box (hereinafter referred to as the partial pressure tank) in the first test, test and use the subsequent test. This procedure does not apply to the partial pressure of the high-pressure tank test.

Document data

Chinese classification (CCS): A55.

Estimated length of the Chinese original: 17,158 words.

Supersedes the earlier edition JJG 531-1988.