

INDUSTRY STANDARD OF THE PEOPLE'S REPUBLIC OF CHINA

JJG 684-2003

Replacing JJG 684-1990

Verification Regulation of Surface Platinum Resistance Thermometer (English Translation)

Issued on: May 12, 2003

Implemented on: November 12, 2003

Issued by: State Administration of Quality Supervision, Inspection and
Quarantine

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to the measurement range of $-60^{\circ}\text{C} \sim 600^{\circ}\text{C}$ (or part of the range) of the contact surface of the platinum RTD's initial verification, testing and use of the follow-up inspection. Surface of the platinum RTD can be made of metal wire around the flat, made of a thin film or thick-film technology, and other processes.

Document data

Chinese classification (CCS): A54.

Estimated length of the Chinese original: 15,169 words.

Supersedes the earlier edition JJG 684-1990.

Standards quoted by this document

JJG 229-1998; JB/T 8622-1997